

Vol 8, Iss 18, Sep 20, 2018 DOI:10.21769/BioProtoc.3015

1

Stable-isotope Labeled Metabolic Analysis in *Drosophila melanogaster*: From Experimental Setup to Data Analysis

Yuping Cai, Nan Liu* and Zheng-Jiang Zhu*

Interdisciplinary Research Center on Biology and Chemistry, Shanghai Institute of Organic Chemistry, Chinese Academy of Sciences, Shanghai, China

*For correspondence: liunan@sioc.ac.cn; jiangzhu@sioc.ac.cn;

[Abstract] Stable-isotope labeled metabolic analysis is an essential methodology to characterize metabolic regulation during biological processes. However, the method using stable-isotope-labeled tracer (e.g., ¹³C-glucose) in live animal is only beginning to be developed. Here, we contribute a qualitative metabolic labeling experiment protocol in *Drosophila melanogaster* using stable-isotope-labeled ¹³C-glucose tracer followed by liquid chromatography-mass spectrometry (LC-MS) analysis. Detailed experimental setup, data acquisition and analysis are provided to facilitate the application of *in vivo* metabolic labeling analysis that might be applied in a wide range of biological studies.

Keywords: Stable-isotope labeling, ¹³C-glucose tracer, Metabolic analysis, Qualitative analysis, Liquid chromatography-mass spectrometry, *Drosophila melanogaster*

[Background] Metabolomics is a newly emergent omic-level study aiming to profile small molecule metabolites in a complex biological system. It has been applied in diverse research areas pertaining to human health and disease, such as biomarker discovery, disease pathogenesis, and assessment of drug toxicity. Measurement of metabolites is important to determine alterations in metabolic pathway in response to endogenous and exogenous changes. To accurately characterize metabolic pathway activity, isotope-labeled tracers (e.g., ¹³C and ¹⁵N) have been used (Park et al., 2016; Jang et al., 2018). There are many such studies (both quantitatively and qualitatively) in cultured cells (Buescher et al., 2015; Liu et al., 2018), however, stable-isotope based metabolic labeling experiment in live animal remain largely unexplored. In the current protocol, we describe a qualitative metabolic labeling analysis by using the labeled ¹³C-glucose as a tracer, and we have successfully applied this protocol to comparatively analyze the activity of glycolysis pathway in *Drosophila melanogaster*, during aging and between wild-type and mutant animals.

Materials and Reagents

A. Consumables

- 1. Pipette tips (Eppendorf, catalog number: 0030073428)
- 2. Ceramic beads (Aoran, catalog number: 150010C)
- 3. Eppendorf tube (2 ml) (Eppendorf, catalog number: 0030120094)
- 4. HPLC glass vial (Agilent Technologies, catalog number: 5182-0716)

@ Copyright Cai et al.



Vol 8, Iss 18, Sep 20, 2018 DOI:10.21769/BioProtoc.3015

- 5. Injection needle (Agilent Technologies, catalog number: G4226-87201)
- 6. Kimwipe filter paper (KCWW, Kimberly-Clark, catalog number: 34120)

B. Biological material

1. Drosophila melanogaster

The *Drosophila* strain used was 5905 (FlyBase ID: FBst0005905, *w*¹¹¹⁸). Flies were cultured in standard media (Recipe 1) at 25 °C with 60% humidity in a 12 h light and 12 h dark cycle. Prior to the test, flies were starved on 1% Agar media for 6 h before transferred to the vials containing a small piece of Kimwipe filter paper (KCWW, Kimberly-Clark, catalog number: 34120) pre-soaked in 1 ml of 10% U-¹³C₆-glucose (U-¹³C₆-glucose was added to phosphate buffer at a final concentration of 10%). Flies were treated for 3 days, and then transferred to new vials with fresh U-¹³C₆-glucose for additional 2 days. Fly heads were dissected from anesthetized flies with CO₂ for subsequent metabolic analysis. For each experiment, 8 biological repeats were conducted, with 20 heads for each repeat. One hundred and sixty male flies were used, with 20 flies per vial.

C. Chemicals

LC-MS chemicals:

- 1. Methanol (MeOH), LC-MS grade (Honeywell, catalog number: LC230-2.5HC). Store at the room temperature (20 °C-25 °C)
- 2. Acetonitrile (ACN), LC-MS grade (Merck, catalog number: 1.00029.2500). Store at the room temperature (20 °C-25 °C)
- 3. Water (H_2O) (Honeywell, catalog number: LC365-2.5HC). Store at the room temperature (20 °C-25 °C)
- 4. Ammonium acetate, LC-MS grade (Sigma-Aldrich, catalog number: 73594-25G-F). Store at 4 °C
- 5. Ammonium hydroxide, LC-MS grade (Sigma-Aldrich, catalog number: 44273-100mL-F). Store at 4 °C
- 6. Liquid nitrogen

Labeled chemicals:

D-Glucose (U-¹³C₆, 99%) (Cambridge Isotope Laboratories, catalog number: <u>CLM-1396-PK</u>).
 Store at the room temperature (20 °C-25 °C)

Drosophila standard media:

- 1. Sucrose
- 2. Maltose
- Yeast
- 4. Agar
- 5. Maizena



Vol 8, Iss 18, Sep 20, 2018 DOI:10.21769/BioProtoc.3015

- 6. Soybean flour
- 7. 438 sodium benzoate
- 8. Methyl-p-hydroxybenzoate
- 9. Propionic acid
- D. Mobile phase setup
 - 1. Mobile phase A (see Recipes)
 - 2. Mobile phase B (see Recipes)

Equipment

- 1. Pipettes
- 2. Homogenizer (BERTIN, model: Precellys® 24)
- 3. Incubator
- 4. Sonicator
- 5. Centrifuge
- 6. Vacuum concentrator (Labconco, German)
- 7. Merck SeQuant ZIC-pHILIC column [particle size, 5 µm; 100 mm (length) x 2.1 mm (i.d.)]
- 8. UHPLC system (Agilent Technologies, model: 1290 Infinity)
- 9. Quadruple time-of-flight mass spectrometer (Agilent Technologies, model: 6550 Series)

Software

- 1. Pathways to PCDL (version B.07.00, Agilent Technologies)
- 2. PCDL Manager (version B.07.00, Agilent Technologies)
- 3. Profinder (version B.08.00, Agilent Technologies)
- 4. MassHunter software (version B.07.00, Agilent Technologies)

Procedure

A. Metabolites extraction

- Quickly freeze the animal tissues (head of *Drosophila*) in liquid nitrogen immediately after dissection.
- 2. Homogenize the tissue sample with 200 μl of H₂O and 5 ceramic beads using the homogenizer.
- 3. Add 800 µl ACN:MeOH (1:1, v/v) to homogenized solution for subsequent metabolite extraction.
- 4. Incubate the samples for 1 h at -20 °C to precipitate proteins.
- 5. Proceed with 15 min centrifugation at 15,000 x g under 4 °C.
- 6. Transfer the resulting supernatant to a new Eppendorf tube (2 ml), then evaporate to dryness in a vacuum concentrator under 4 °C.



Vol 8, Iss 18, Sep 20, 2018 DOI:10.21769/BioProtoc.3015

- 7. Reconstitute the dry extracts with 100 μl of ACN:H₂O (1:1, v/v).
- 8. Sonicate the reconstitution solution for 10 min, and centrifuge for 15 min at 15,000 *x g* under 4 °C to remove insoluble debris.
- 9. Transfer the supernatant to an HPLC glass vial and store at -80 °C if the samples will be subjected to LC-MS analysis within 3 h. For extracted samples that require long time (over 12 h) stored prior to being analyzed, we suggest storing the samples after Step A6 and then proceeding with A8-A9 before LC-MS analysis.

B. LC-MS analysis

- 1. Liquid chromatography
 - a. Load worklist with method embedded using MassHunter software. Please note that LC-MS operation (both instrument and software) requires specialized training.
 - b. Run batch sequence with following LC parameters:
 - i. Wash injection needle one time with needle washing solvent MeOH:H2O (1:1, v/v).
 - ii. Load sample and inject 2 μl of sample.
 - iii. Run LC method using the LC gradient as described in Table 1.

Table 1. The gradient elution method for LC-MS analysis

Time (min)	Flow rate (ml/min)	Eluent A (vol. %)	Eluent B (vol. %)
0.0	0.2	20.0	80.0
2.0	0.2	20.0	80.0
17.0	0.2	80.0	20.0
17.1	0.2	20.0	80.0
22.1	0.4	20.0	80.0
22.2	0.2	20.0	80.0

2. Mass spectrometry

Set MS parameters as described below:

- a. ESI source parameters:
 - Sheath gas temperature, 300 °C.
 - ii. Dry gas temperature, 250 °C.
 - iii. Sheath gas flow, 12 L/min.
 - iv. Dry gas flow, 16 L/min.
 - v. Capillary voltage, 2,500 V (+) and -2,500 V (-), respectively. Please note that the same sample is analyzed twice for each ionization mode.
 - vi. Nozzle voltage, 0 V.
 - vii. Nebulizer pressure, 20 psi.
- b. Time of Flight (TOF) parameters:
 - i. TOF scan range: m/z 60-1,200 Da.

Vol 8, Iss 18, Sep 20, 2018 DOI:10.21769/BioProtoc.3015

ii. MS1 acquisition frequency: 4 Hz.

Data analysis

A. Extraction of isotopologues

1. Metabolite library construction

Use Pathways to PCDL software (version B.07.00, Agilent Technologies) and PCDL Manager software (version B.07.00, Agilent Technologies) to build a metabolite library for metabolites in both glycolysis and citric acid cycle. Specifically, each metabolite standard is analyzed under the same LC-MS condition as biological samples. The ion chromatograph of each metabolite is extracted to obtain the retention time information. Then, the retention time together with formula value is used to construct a metabolite library using PCDL manager. The input example is provided as below (Table 2):

Table 2. The metabolite library for metabolites in both glycolysis and citric acid cycle

Metabolite name	Formula	Retention time (min)
Glucose	C ₆ H ₁₁ NaO ₆	5.1
F6P	$C_6H_{13}O_9P$	7.8
G6P	$C_6H_{13}O_9P$	8.2
GADP	$C_3H_7O_6P$	7.4
PEP	$C_3H_5O_6P$	8.7
Pyruvate	$C_3H_4O_3$	7.3
Lactate	$C_3H_6O_3$	3.2
Citrate	$C_6H_8O_7$	8.9
Malate	$C_4H_6O_5$	7.8
Fumarate	$C_4H_4O_4$	7.8
Cis-aconitate	$C_6H_6O_6$	8.9
α-KG	$C_5H_6O_5$	7.5
Succinate	C ₄ H ₆ O ₄	7.4

2. Raw data loading

Load the acquired LC-MS raw data files (.d) into Profinder (version B.08.00, Agilent Technologies) for the extraction of metabolite isotopologues using the constructed metabolite library.

- 3. Feature extraction parameters:
 - a. Ion abundance criterion: peak core area 20% of peak height.
 - b. Mass tolerance: ± 15 ppm + 2.00 mDa.
 - c. Retention time tolerance: ± 0.20 min.
 - d. Anchor ion height threshold: 250 counts.



Vol 8, Iss 18, Sep 20, 2018 DOI:10.21769/BioProtoc.3015

e. Sum of ion heights threshold: 1,000 counts.

f. Correlation coefficient threshold: 0.5.

B. Stable-isotope-labeled metabolic analysis

1. Peak integration result manual check

After isotopologues extraction in Profinder, peak integration result need to be reviewed and manually curated for subsequent accurate stable-isotope labeled metabolic analysis. Make sure the peak integration range is consistent across multiple samples. Figure 1A illustrates the extracted ion chromatography (EIC) of the key metabolite $^{13}C_3$ -lactate and the peak integration range.

2. Calculation of tracer incorporation

For each targeted metabolite, different isotope pattern will be obtained corresponding to the number of incorporated 13 C atoms. For example, the isotopologues of lactate are m + 0, m + 1, m + 2, and m + 3 (Figure 1B).

a. Abundance of individual isotopologue is the integrated peak area.

Taken metabolite lactate as an example, Figure 1C shows the abundance level of one isotopologue of lactate M + 3 between two groups (wild type and PRC2 mutant). In the article by Ma *et al.*, 2018, Figure 6G is generated using this calculated data.

b. Total metabolite abundance

Total metabolite abundance is calculated using the following formula:

total abundance = intensity of
$$\sum ([M0 + M1 + M2 + \cdots Mn])$$
 Eq. 1

Mn is the labeling pattern of the isotopologue with all atoms (C or N) labeled.

Figure 1D shows the total abundance level of lactate between two groups (wild type and PRC2 mutant). In the article by Ma *et al.*, 2018, Figure S5E was generated using this calculated data.

c. Proportion of individual isotopologue

Proportion of individual isotopologue is calculated using the following formula:

% of total pool = intensity of
$$\frac{[Mi]}{\sum([M0 + M1 + M2 + \cdots Mn])} \times 100\%$$
 Eq. 2

Mi is the labeling pattern of individual isotopologue.

Mn is the labeling pattern of the isotopologue with all atoms (C or N) labeled.

Figure 1E shows the percentage of total pool level of one isotopologue of lactate M + 3 between two groups (wild type and PRC2 mutant). In the article by Ma *et al.*, 2018, Figure S5B and Figure S5D were generated using this calculated data.

d. Total tracer incorporation



Vol 8, Iss 18, Sep 20, 2018 DOI:10.21769/BioProtoc.3015

Total tracer incorporation is calculated using the following formula:

total tracer incorporation (%) = intensity of
$$\frac{\sum([M1 + M2 + \cdots Mn])}{\sum([M0 + M1 + M2 + \cdots Mn])} \times 100\%$$
 Eq. 3

Figure 1F shows the percentage of total tracer incorporation between two groups (wild type and PRC2 mutant).

Above results demonstrated that lactate, the end product of glycolysis pathway, significantly increased in PRC2 mutants.




Figure 1. Stable-isotope-labeled metabolic analysis strategy. A. The extracted ion chromatography (EIC) of isotopologue 13 C₃-lactate (m + 3). (mean ± SD of 8 biological repeats with 10 flies for each measurement; Student's *t*-test; n.s.: not significant). Test was from muscle tissues of 30 d old male flies. Genotypes: WT: 5905. Mut: Pcl^{c421} /+; $Su(z)12^{c253}$ /+. B. The labeling pattern of lactate demonstrated in mass spectrum (m + 0, m + 1, m + 2, and m + 3). C. The abundance level of lactate isotopologue m + 3 between two groups (mean ± SD of 8 biological repeats with 10 flies for each measurement; Wilcox test). Genotypes: WT: 5905. Mut: Pcl^{c421} /+; $Su(z)12^{c253}$ /+. D. The total abundance level of lactate between two groups (wild type and PRC2 mutant) (mean ± SD of 8 biological repeats with 10 flies for each measurement; Wilcox test). Genotypes: WT: 5905. Mut: Pcl^{c421} /+; $Su(z)12^{c253}$ /+. E. The percentage of total pool level of one isotopologue of lactate m + 3 between two groups (mean ± SD of 8 biological repeats with 10 flies for each measurement; Wilcox test). Genotypes: WT: 5905. Mut: Pcl^{c421} /+; $Su(z)12^{c253}$ /+. F. The percentage of total tracer incorporation between two groups (mean ± SD of 8 biological repeats with 10 flies for each measurement; Wilcox test). Genotypes: WT: 5905. Mut: Pcl^{c421} /+; $Su(z)12^{c253}$ /+.

© Copyright Cai et al.



Vol 8, Iss 18, Sep 20, 2018 DOI:10.21769/BioProtoc.3015

Recipes

1. Standard Drosophila food

Sucrose 36 q/L

Maltose 38 g/L

Yeast 22.5 g/L

Agar 5.4 g/L

Maizena 60 g/L

Soybean flour 8.25 g/L

438 sodium benzoate 0.9 g/L

Methyl-p-hydroxybenzoate 0.225 g/L

Propionic acid 6.18 ml/L

ddH₂O to make up 1 L

2. Mobile phase A

25 mM ammonium acetate

25 mM ammonium hydroxide

For the preparation of 1 L mobile phase A, firstly weigh 1.9271 g CH_3COONH_4 . Dissolve the CH_3COONH_4 in 1 L H_2O . Then add 3.5 ml NH_4OH (25%) to generate the mobile phase A. Store the solution at 4 °C for up to 2 weeks

3. Mobile phase B

Acetonitrile

Store at the room temperature (20-25 °C)

Acknowledgments

We thank the financial support provided by the startup funding from Interdisciplinary Research Center on Biology and Chemistry (IRCBC), and Agilent Technologies Thought Leader Award. N.L. and Z.-J. Z. are also supported by Thousand Youth Talents Program. This protocol is also a part of our previous work by Ma *et al.*, 2018.

Competing interests

The authors declare no competing financial interest.

References

Buescher, J. M., Antoniewicz, M. R., Boros, L. G., Burgess, S. C., Brunengraber, H., Clish, C. B., DeBerardinis, R. J., Feron, O., Frezza, C., Ghesquiere, B., Gottlieb, E., Hiller, K., Jones, R. G., Kamphorst, J. J., Kibbey, R. G., Kimmelman, A. C., Locasale, J. W., Lunt, S. Y., Maddocks,

© Copyright Cai et al.



Vol 8, Iss 18, Sep 20, 2018 DOI:10.21769/BioProtoc.3015

- O. D., Malloy, C., Metallo, C. M., Meuillet, E. J., Munger, J., Noh, K., Rabinowitz, J. D., Ralser, M., Sauer, U., Stephanopoulos, G., St-Pierre, J., Tennant, D. A., Wittmann, C., Vander Heiden, M. G., Vazquez, A., Vousden, K., Young, J. D., Zamboni, N. and Fendt, S. M. (2015). <u>A roadmap</u> for interpreting ¹³C metabolite labeling patterns from cells. *Curr Opin Biotechnol* 34: 189-201.
- 2. Jang, C., Chen, L. and Rabinowitz, J. D. (2018). <u>Metabolomics and isotope tracing.</u> *Cell* 173(4): 822-837.
- Liu, L., Su, X., Quinn, W. J., 3rd, Hui, S., Krukenberg, K., Frederick, D. W., Redpath, P., Zhan, L., Chellappa, K., White, E., Migaud, M., Mitchison, T. J., Baur, J. A. and Rabinowitz, J. D. (2018). Quantitative analysis of NAD synthesis-breakdown fluxes. Cell Metab 27(5): 1067-1080 e1065.
- 4. Ma, Z., Wang, H., Cai, Y., Wang, H., Niu, K., Wu, X., Ma, H., Yang, Y., Tong, W., Liu, F., Liu, Z., Zhang, Y., Liu, R., Zhu, Z. J. and Liu, N. (2018). Epigenetic drift of H3K27me3 in aging links glycolysis to healthy longevity in *Drosophila*. *Elife* 7: e35368.
- Park, J. O., Rubin, S. A., Xu, Y. F., Amador-Noguez, D., Fan, J., Shlomi, T. and Rabinowitz, J. D. (2016). <u>Metabolite concentrations</u>, <u>fluxes and free energies imply efficient enzyme usage</u>. *Nat Chem Biol* 12(7): 482-489.